

Technical evaluation of bids against our tender call notice NIT-16/2018(B-000041-BIO-17-18) for purchase of steady state fluorimeter with accessories.

Against our tender call notice two bids are received:

1. Base model QM-8075-11-C from Horiba Instruments Incorporated, New Jersey, USA. - through their dealer Analytical Instruments Sales and Service Pvt. Ltd. Kolkata
2. Base model FLS1000-SS-s from Edinburgh Instruments Ltd., Livingstone, UK - through their dealer Specialise Instruments Marketing Company, Mumbai


A technical committee constituted by Director, NISER, met on 24<sup>th</sup> September, 2018 to evaluate the bids for compliance and other technical aspects of the both the models.


Technical committee sought clarifications from both the vendors, where ever the information provided in the bids was not sufficient to evaluate the bids.


Technical committee is satisfied with the clarifications received from both the vendors.


Both the quoted models comply with the required technical specifications as per the tender notice, except that the model FLS1000-SS-s of Edinburgh Instruments Ltd., quoted by Specialise Instruments Marketing Company, Mumbai cannot perform a simultaneous UV-VIS and fluorescence measurements. However, technical committee is satisfied with the alternative mode provided for UV-VIS and fluorescence measurements in the Edinburgh Instruments model.


Technical committee recommends both the bids for financial comparison.

  
Dr. Moloy Sarkar

  
Dr. Chandan Goswami 5/10/18

  
Dr. Pankaj V. Alone

  
Dr. Ramanujam Srinivasan

  
Dr. Tirumala Kumar Chowdary 5.10.18