




Technical compliance report for the E-Tender No. NC-000017-BIO-18-19

S.No.	Item Specification	Labeast Instruments	Firstsource laboratory solutions LLP
LIQUID NITROGEN DEWAR WITH DISPENSER			
1.	Liquid Nitrogen Capacity - 30 Litres - 35 Litres, preferable 35 Litres	Complied	Complied
2.	Static evaporation rate L/day - 0.22	Complied	Complied
3.	Neck diameter - 64mm	Complied	Complied
4.	Usable height - 378mm	Complied	Complied
5.	Outer diameter - 432mm	Complied	Complied
6.	Internal diameter - 356mm	Complied	Complied
7.	Empty weight - 11.4kg	Complied	Complied
TRANSFER UNIT			
1.	Liquid nitrogen capacity - 2.5L	Complied	Complied
2.	Neck diameter - 149mm	Complied	Complied
3.	Usable height - 143mm	Complied	Complied
4.	Outer diameter - 203mm	Complied	Complied
5.	Internal diameter - 152mm	Complied	Complied
6.	Empty weight - 3.6kg	Complied	Complied
7.	Warranty - 1 year	Complied	Complied

The technical evaluation committee comprising Dr. K.C. Panigrahi, Dr. B. Chandramohan and Chairperson, SBS compared the technical bid received from two companies- Labeast Instruments and Firstsource laboratory solutions LLP and found both the companies to be technically qualified.


 Dr. K.C. Panigrahi
 Indentor


 Dr. B. Chandramohan
 Member


 Chairperson
 SBS
 21/5/19