


Subject: Technical Comparative Statement for purchase Double Stage Diaphragm Vacuum Pump

E-Tender No. NC-000947-CHE-18-19-A

SNo	Specification for Double Stage Diaphragm Vacuum Pump	Microlabs analytical associates	Aditya scientific
1	Chemical-resistant, 2 stage PTFE diaphragm for use in the laboratory flow rate of 2.0 m3/h and an ultimate vacuum of 9 mbar.	Complied	Complied
2	Ultimate vacuum – 9 mbar	Complied	Complied
3	Power Input – 180 watts	Complied	Complied
4	Max back pressure – 1 bar	Complied	Complied
5	Max pumping speed – 2.0 m3/h	Complied	Complied
6	Inlet Connection – 10mm	Complied	Complied
7	Outlet Connection – 10mm	Complied	Complied
8	Noise level – 45 Dba	Complied	Complied
9	Supply voltage: 230 VAC, 50 HZ.	Complied	Complied
10	03 year warranty	Complied	Complied

* Aditya scientific and Microlabs analytical associates met all the requirements and specifications therefore can be considered for further price comparison.


Intendor


10/07/2019
Chairperson