


**Technical Comparison for tender No. NC-001173-NC-001173-12-R&D-
NIS-5.03-0100-18-19 for Magnetic annealing system**

Company Name		EXCEL INSTRUMENTS
S/N	Specification	Comply.?
1	Vacuum compatible small dia heating arrangement for magnetic annealing for sample size 10mmX10mm (Max).	Yes
2	Max temperature 500C	Yes
3	K type thermocouple,	Yes
4	PID programmable temperature controller.	Yes
5	Connections for vacuum pump.	Yes
6	Right angle isolation valve HSN 8479	Yes
7	01 year free onsite comprehensive warranty from the date of installation	Yes

Recommendation: With comparison of technical specifications supplied by one vendors, we concluded that Excel Instruments is fulfilling all requirements. Please open the financial bid of the same.


16/08/19.
Dr. Subhankar Bedanta
Indenting Officer